



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description: ZYMUTEST vWF: CBA

Product Code: RK038A

Lot No.: FB0628

Expiration date: 2025-05-12

Components	Quantity	Lot	Expiration Date
Collagen pre-coated plate	12x8 wells	FB0704	2025-09-30
vWF standard	3 vials	FB0967	2025-11-20
vWF Control I	1 vial	FB0700	2025-10-02
vWF Control II	1 vial	FB0701	2025-10-02
Anti-(h)- vWF HRP immunoconjugate	3 vials	FB0731	2025-10-06
Sample diluent	2x50 ml	FB0233	2025-07-14
Wash solution	1x50 ml	FA2213	2025-05-12
Conjugate diluent	1x25 ml	FB0595	2025-09-15
TMB	1x25 ml	210604D01	2025-06-30
Sulfuric Acid 0,45M	1x6 ml	FB0237	2025-07-16

Analytical Data	Specification	Result
Reactivity	A450 for 100 % ≥ 1.00	2.28
Calibration Curve and Accuracy	Controls Measured Values within acceptance ranges	PASS
Blank	< 0.100	< 0.100
Detection threshold	$\leq 5\%$	$< 1\%$

Calibration Curve for information

Conc.	C	C/2	C/4	C/10	C/20
OD	2.582	1.822	1.147	0.532	0.292

	Acceptance Range	MV
CI	[78 - 106]	87
CII	[39 - 58]	44

MV = Measured value

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval: 2022-06-28

Quality Control Manager: S.LECOURT

