

A Company of Sysmex Group



HYPHEN BioMed 155 rue d'Eragny 95000 Neuville-sur-Oise France

CERTIFICATE OF ANALYSIS

Product Description:

BIOPHEN HEPARIN CALIBRATOR

Product Code:

222001

Lot No.:

FC00161F

Expiration date :

2025-06-24

Components	Quantity	Lot	Expiration Date
CAL 1	4 vials	FC0016BJ1F	2025-06-24
CAL 2	4 vials	FC0016BK1F	2025-06-24
CAL 3	4 vials	FC0016BL1F	2025-06-24
CAL 4	4 vials	FC0016BM1F	2025-06-24
CAL 5	4 vials	FC0016BN1F	2025-06-24

Analytical Data	Specification	Result
Concentration	CAL 1 : ≤ 0.05 IU/mI CAL 2 : 0.30 -0.50 IU/mI CAL 3 : 0.70 -0.90 IU/mI CAL 4 : 1.10 -1.30 IU/mI CAL 5 : 1.45 -1.75 IU/mI	0 IU/mL 0.37 IU/mL 0.77 IU/mL 1.14 IU/mL 1.51 IU/mL
Linearity and accuracy of calibration with values assigned	$R^2 \ge 0.98$ Measured values for controls in acceptance ranges	PASS

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval: 2023-01-19

Quality Control Manager:

S.LECOURT