



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : ZYMUTEST PAI-1 Activity

Product Code : RK019A

Lot No. : FB1795

Expiration date : 2025-06-30

Components	Quantity	Lot	Expiration Date
tPA pre-coated plate	12x8 wells	FB1823	2026-04-20
(h) PAI-1 standard	3 vials	FB1798	2026-04-15
(h) PAI-1 Control I	1 vial	FB1800	2026-04-15
(h) PAI-1 Control II	1 vial	FB1801	2026-04-15
Anti-(h)-PAI-1 HRP immunoconjugate	3 vials	FB1851	2026-04-27
F-Sample diluent	2x50 ml	FB1825	2026-04-20
Wash solution	1x50 ml	FB0867	2025-11-03
Conjugate diluent	1x25 ml	FB1457	2026-02-26
TMB	1x25 ml	210604D01	2025-06-30
Sulfuric Acid 0,45M	1x6 ml	FB1450	2026-02-23

Analytical Data	Specification	Result
Reactivity	A450 for 8 ng/ml $\geq$ 1.00	1.979
Calibration Curve and Accuracy	Controls Measured Values within acceptance ranges	PASS
Blank	< 0.100	< 0.01
Detection threshold	$\leq$ 0.5 ng/ml	< 0.1 ng/mL

Calibration Curve for information

Conc.	C	C/2	C/4	C/10	C/20
OD	2.200	1.430	0.852	0.363	0.197

	Acceptance Range	MV
CI	[8.1 - 11.0]	9.1
CII	[3.0 - 4.4]	3.7

MV = Measured value

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2022-12-02

Quality Control Manager : S. LECOURT