



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : ZYMUTEST FACTOR X

Product Code : RK033A

Lot No. : FB1575

Expiration date : 2025-06-30

Components	Quantity	Lot	Expiration Date
Anti (h) Factor X pre-coated plate	12x8 wells	FB1576	2026-03-16
Factor X standard	3 vials	FB1531	2026-03-12
Factor X Control I	1 vial	FB1533	2026-03-12
Factor X Control II	1 vial	FB1532	2026-03-12
Anti-(h) Factor X HRP immunoconjugate	3 vials	FB1577	2026-03-18
Sample diluent	2x50 ml	FB0775	2025-10-20
Wash solution	1x50 ml	FB0867	2025-11-03
Conjugate diluent	1x25 ml	FB1457	2026-02-26
TMB	1x25 ml	210604D01	2025-06-30
Sulfuric Acid 0,45M	1x6 ml	FB1450	2026-02-23

Analytical Data	Specification	Result
Reactivity	A450 for 100 % ≥ 1.40	1.57
Calibration Curve and Accuracy	Controls Measured Values within acceptance ranges	PASS
Blank	< 0.100	0.021
Detection threshold	$\leq 5\%$	$< 1\%$

Calibration Curve for information

Conc.	C	C/2	C/4	C/10	C/20
OD	1.670	1.203	0.760	0.369	0.217

	Acceptance Range	MV
CI	[65-89]	80
CII	[23-35]	28

MV = Measured value

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2022-10-18

Quality Control Manager : S.LECOURT