



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : ZYMUTEST vWF: CBA

Product Code : RK038A

Lot No. : FB1306

Expiration date : 2025-06-30

Components	Quantity	Lot	Expiration Date
Collagen pre-coated plate	12x8 wells	FB1409	2026-02-09
vWF standard	3 vials	FB1318	2026-01-29
vWF Control I	1 vial	FB1316	2026-01-29
vWF Control II	1 vial	FB1317	2026-01-29
Anti-(h)- vWF HRP immunoconjugate	3 vials	FB1411	2026-02-12
Sample diluent	2x50 ml	FB0774	2025-10-20
Wash solution	1x50 ml	FB0445	2025-08-18
Conjugate diluent	1x25 ml	FB0764	2025-10-16
TMB	1x25 ml	210604D01	2025-06-30
Sulfuric Acid 0,45M	1x6 ml	FB0237	2025-07-16

Analytical Data	Specification	Result
Reactivity	A450 for 100 % ≥ 1.00	1.59
Calibration Curve and Accuracy	Controls Measured Values within acceptance ranges	PASS
Blank	< 0.100	0.006
Detection threshold	$\leq 5\%$	$< 1\%$

Calibration Curve for information

Conc.	C	C/2	C/4	C/10	C/20
OD	1.733	1.214	0.769	0.352	0.190

	Acceptance Range	MV
CI	[70 - 95]	84
CII	[35 - 53]	42

MV = Measured value

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 22/09/2022

Quality Control Manager :

