

HYPHEN BioMed 155 rue d'Eragny 95000 Neuville-sur-Oise France

CERTIFICATE OF ANALYSIS

Item Description: LIAPHEN™ Protein C

Item Code: 120003

Lot No.: FC255016

Expiration date: 2025-07-01

Components	Quantity	Lot	Expiration Date
R1: Reaction Buffer	4 vials	FC2550en16	2025-07-01
R2: Latex reagent	4 vials	FC2550eo16	2025-07-01

Analytical Data	Specification	Result
Calibration curve : reactivity, slope and accuracy	OD/min $(3/2) \ge 0.30$ Delta OD/min $(1/1-1/2) \ge 0.1$ Controls measured within acceptance ranges	0.43 0.1 PASS
Hook effect	OD/min (300 et 600 %) ≥ OD/min (point 1/1 of calibration curve)	PASS

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval:

Quality Control Approval Quality Control Manager:

13 FEV. 2024

S. LECOURT