



A Company of Sysmex Group



HYPHEN BioMed  
155 rue d'Eragny  
95000 Neuville-sur-Oise  
France

## CERTIFICATE OF ANALYSIS

**Product Description :** BIOPHEN LMWH CONTROL PLASMA  
**Product Code :** 223001  
**Lot No. :** FC2162  
**Expiration date :** 2025-07-19

| Components | Quantity | Lot        | Expiration Date |
|------------|----------|------------|-----------------|
| C3         | 6 vials  | FC1882BO2K | 2026-02-14      |
| C4         | 6 vials  | FC0229BP2E | 2025-07-19      |

| Analytical Data | Specification                                  | Result                   |
|-----------------|--|--------------------------|
| Concentration   | C3 : 0.70 -0.90 IU/ml<br>C4 : 1.05 -1.35 IU/ml | 0.77 IU/mL<br>1.17 IU/mL |

*This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.*

Date of Approval : 2023-10-05

Quality Control Manager : S.LECOURT

P/O