

A Company of Sysmex Group



HYPHEN BioMed 155 rue d'Eragny 95000 Neuville-sur-Oise France

CERTIFICATE OF ANALYSIS

Product Description:

BIOPHEN UFH CONTROL PLASMA

Product Code:

223101

Lot No.:

FC1428

Expiration date:

2025-07-14

Components	Quantity	Lot	Expiration Date
C1	6 vials	FC0196BX1D	2025-07-14
C2	6 vials	FC0943BY1D	2025-09-06

Analytical Data	Specification	Result
Concentration	C1 : 0.10 -0.30 IU/ml C2 : 0.40 -0.60 IU/ml	0.21 IU/lmL 0.50 IU/lmL

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval: 2025-05-31

Quality Control Manager:

S.LECOURT