



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN LMWH CONTROL LOW
Product Code : 223701
Lot No. : FC2336
Expiration date : 2025-07-22

Components	Quantity	Lot	Expiration Date
CI	6 vials	FC2112BQG6	2026-03-21
CII	6 vials	FC0233BR1E	2025-07-22

Analytical Data	Specification	Result
Concentration	CI : 0.15 -0.35 IU/ml CII : 0.35 -0.65 IU/ml	0.30IU/mL 0.46IU/mL

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval :2023-11-16

Quality Control Manager :S.LECOURT