

 HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : ADP
Product Code : AG001K
Lot No. : FC2174
Expiration date : 2025-07-29

Components	Quantity	Lot	Expiration Date
ADP	3 vials	FC2174	2025-07-29

Analytical Data	Specification	Result
Reactivity (% AgregMax) at 2µM on CS or CN series	≥ 70%	83%
Reactivity (% AgregMax) at 2µM on agregometer	≥ 70%	72%

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval :2023-10-16

Quality Control Manager : S.LECOURT

