



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : RISTOCETIN
Product Code : AG004K
Lot No. : FC0678
Expiration date : 2025-08-09

Components	Quantity	Lot	Expiration Date
RISTOCETIN	3 vials	FC0678	2025-08-09

Analytical Data	Specification	Result
Reactivity (% AgregMax) at 1.2mg/ml on CS	≥ 50%	81%
Reactivity (% AgregMax) at 1.2mg/ml on Agregometer	≥ 50%	78%

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2023-03-20

Quality Control Manager : **S. LECOURT**

