

 HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : IMIDAZOLE BUFFER
Product Code : AR021K/L
Lot No. : FC1167E7
Expiration date : 2025-10-10

Components	Quantity	Lot	Expiration Date
IMIDAZOLE BUFFER	3 vials *25mL (AR021K) 12 vials *25mL (AR021L)	FC1167ARE7	2025-10-10

Analytical Data	Specification	Result
Hemoclot PC clotting assay	Delta CT (100%) with control sample \leq 5%	1.5%
FVIII:C clotting assay	Delta CT (100%) with control sample \leq 5%	1.1%

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2023-05-09

Quality Control Manager :

S.LECOURT

