



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : ZYMUTEST vWF: CBA

Product Code : RK038A

Lot No. : FC0004

Expiration date : 2025-10-16

Components	Quantity	Lot	Expiration Date
Collagen pre-coated plate	12x8 wells	FB2052	2026-05-25
vWF standard	3 vials	FC0021	2026-06-16
vWF Control I	1 vial	FC0019	2026-06-16
vWF Control II	1 vial	FC0020	2026-06-16
Anti-(h)- vWF HRP immunoconjugate	3 vials	FC0068	2026-06-25
Sample diluent	2x50 ml	FB1933	2026-05-11
Wash solution	1x50 ml	FB0866	2025-11-03
Conjugate diluent	1x25 ml	FB0764	2025-10-16
TMB	1x25 ml	220106D03	2026-01-31
Sulfuric Acid 0,45M	1x6 ml	FB1449	2026-02-23

Analytical Data	Specification	Result
Reactivity	A450 for 100 % ≥ 1.00	1.86
Calibration Curve and Accuracy	Controls Measured Values within acceptance ranges	PASS
Blank	< 0.100	0.012
Detection threshold	$\leq 5\%$	$< 1\%$

Calibration Curve for information

Conc.	C	C/2	C/4	C/10	C/20
OD	2.155	1.545	1.016	0.491	0.269

	Acceptance Range	MV
CI	[65 - 87]	74
CII	[32 - 48]	38


MV = Measured value

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2023-01-25

Quality Control Manager :

S.LECOURT



F820-11/ZY/038A v10