



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : ZYMUTEST vWF: CBA

Product Code : RK038A

Lot No. : FC0712

Expiration date : 2026-01-31

Components	Quantity	Lot	Expiration Date
Collagen pre-coated plate	12x8 wells	FC1045	2026-09-15
vWF standard	3 vials	FC0895	2026-08-24
vWF Control I	1 vial	FC0893	2026-08-24
vWF Control II	1 vial	FC0894	2026-08-24
Anti-(h)- vWF HRP immunoconjugate	3 vials	FC0928	2026-09-03
Sample diluent	2x50 ml	FB1933	2026-05-11
Wash solution	1x50 ml	FC0034	2026-06-15
Conjugate diluent	1x25 ml	FC0128	2026-06-30
TMB	1x25 ml	220106D03	2026-01-31
Sulfuric Acid 0,45M	1x6 ml	FC0223	2026-07-15

Analytical Data	Specification	Result
Reactivity	A450 for 100 % ≥ 1.00	1.86
Calibration Curve and Accuracy	Controls Measured Values within acceptance ranges	PASS
Blank	< 0.100	0.0001
Detection threshold	$\leq 5\%$	$< 1\%$

Calibration Curve for information

Conc.	C	C/2	C/4	C/10	C/20
OD	2.01	1.49	0.93	0.44	0.24


	Acceptance Range	MV
CI	[59-79]	66
CII	[38 -56]	48

MV = Measured value

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2023-05-16

Quality Control Manager : S.LECOURT



F820-11/ZY/038A v10