



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : ADP
Product Code : AG001K
Lot No. : FC2188
Expiration date : 2026-04-03

Components	Quantity	Lot	Expiration Date
ADP	3 vials	FC2188	2026-04-03

Analytical Data	Specification	Result
Reactivity (% AgregMax) at 2 μ M on CS or CN series	$\geq 70\%$	86%
Reactivity (% AgregMax) at 2 μ M on agregometer	$\geq 70\%$	85%

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2023-11-02

Quality Control Manager : S.LECOURT

